



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040723\
Data File : VN077266.D
Acq On : 07 Apr 2023 08:38
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N040723W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Fri Apr 07 13:04:49 2023



AutoFind: Scans 1855, 1856, 1857; Background Corrected with Scan 1847

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.6	23440	PASS
75	95	30	60	52.8	70365	PASS
95	95	100	100	100.0	133339	PASS
96	95	5	9	6.4	8513	PASS
173	174	0.00	2	0.7	718	PASS
174	95	50	100	73.6	98131	PASS
175	174	5	9	7.5	7322	PASS
176	174	95	101	97.8	95976	PASS
177	176	5	9	6.5	6199	PASS