

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040820\
 Data File : VN060907.D
 Acq On : 8 Apr 2020 17:33
 Operator : JC/MD
 Sample : L2185-01 20X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 30773

Quant Time: Apr 09 09:20:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	195935	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	326058	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	311111	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	146527	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	127561	47.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.86%	
35) Dibromofluoromethane	7.56	113	93898	47.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.26%	
50) Toluene-d8	10.08	98	405498	50.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.70%	
62) 4-Bromofluorobenzene	12.40	95	155023	52.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.06%	
Target Compounds						
16) Acetone	3.77	43	4344505	4559.907	ug/l	95
52) Toluene	10.14	92	926854	161.022	ug/l	99
68) m/p-Xylenes	11.61	106	9451	2.217	ug/l	97
69) o-Xylene	11.94	106	5826	1.432	ug/l	99
84) 1,2,4-Trimethylbenzene	13.04	105	30544	3.309	ug/l	100
95) Naphthalene	15.13	128	18243	2.210	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040820\
Data File : VN060907.D
Acq On : 8 Apr 2020 17:33
Operator : JC/MD
Sample : L2185-01 20X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
30773

Quant Time: Apr 09 09:20:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

