

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040821\
 Data File : VN066535.D
 Acq On : 8 Apr 2021 12:55
 Operator : JC/MD
 Sample : M1961-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR1-A-BOTTOM

Manual Integrations
 APPROVED

MMDadoda
 4/9/2021 1:53:51 PM

Quant Time: Apr 09 04:33:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 06 15:38:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.587	168	307484	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	463555	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	444972	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	180895	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.947	65	201715	49.81	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	99.62%		
35) Dibromofluoromethane	7.507	113	158924	50.30	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	100.60%		
50) Toluene-d8	10.028	98	596048	49.65	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.30%		
62) 4-Bromofluorobenzene	12.348	95	177810	45.68	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	91.36%		
Target Compounds						Qvalue
16) Acetone	3.741	43	25865	15.24	ug/l	97
18) Methyl Acetate	4.237	43	4345m	1.02	ug/l	
20) Methylene Chloride	4.455	84	19142	5.25	ug/l	91
43) Isopropyl Acetate	8.089	43	28470	3.30	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040821\
Data File : VN066535.D
Acq On : 8 Apr 2021 12:55
Operator : JC/MD
Sample : M1961-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR1-A-BOTTOM

Manual Integrations
APPROVED

MMDadoda
4/9/2021 1:53:51 PM

Quant Time: Apr 09 04:33:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 06 15:38:34 2021
Response via : Initial Calibration

