

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040821\
 Data File : VN066537.D
 Acq On : 8 Apr 2021 13:41
 Operator : JC/MD
 Sample : M1961-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR1-B-BOTTOM

Quant Time: Apr 09 04:33:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 06 15:38:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.584	168	311655	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	475549	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	461782	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	191885	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.946	65	203098	49.48	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.96%
35) Dibromofluoromethane	7.509	113	161741	49.90	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.80%
50) Toluene-d8	10.028	98	606251	49.23	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.46%
62) 4-Bromofluorobenzene	12.345	95	177872	44.54	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.08%
Target Compounds						
					Qvalue	
16) Acetone	3.738	43	25933	15.07	ug/l	100
20) Methylene Chloride	4.463	84	19549	5.29	ug/l	91
43) Isopropyl Acetate	8.086	43	27756	3.13	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040821\
Data File : VN066537.D
Acq On : 8 Apr 2021 13:41
Operator : JC/MD
Sample : M1961-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR1-B-BOTTOM

Quant Time: Apr 09 04:33:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 06 15:38:34 2021
Response via : Initial Calibration

