

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040821\
 Data File : VN066543.D
 Acq On : 8 Apr 2021 15:59
 Operator : JC/MD
 Sample : M1961-16
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR2-A-BOTTOM

Manual Integrations
 APPROVED

MMDadoda
 4/9/2021 1:54:41 PM

Quant Time: Apr 09 04:35:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 06 15:38:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.587	168	308686	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	469139	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	461105	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	190271	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.946	65	198710	48.88	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	97.76%		
35) Dibromofluoromethane	7.507	113	159292	49.82	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.64%		
50) Toluene-d8	10.028	98	605418	49.83	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.66%		
62) 4-Bromofluorobenzene	12.348	95	183022	46.46	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	92.92%		
Target Compounds						Qvalue
16) Acetone	3.738	43	22156	13.00	ug/l	98
18) Methyl Acetate	4.243	43	4397m	1.03	ug/l	
20) Methylene Chloride	4.460	84	20048	5.48	ug/l	88
43) Isopropyl Acetate	8.089	43	27051	3.10	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040821\
Data File : VN066543.D
Acq On : 8 Apr 2021 15:59
Operator : JC/MD
Sample : M1961-16
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR2-A-BOTTOM

Manual Integrations
APPROVED

MMDadoda
4/9/2021 1:54:41 PM

Quant Time: Apr 09 04:35:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 06 15:38:34 2021
Response via : Initial Calibration

