

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040821\
 Data File : VN066547.D
 Acq On : 9 Apr 2021 1:19
 Operator : JC/MD
 Sample : M1961-20
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR2-C-BOTTOM

Quant Time: Apr 09 04:36:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 06 15:38:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.582	168	299678	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.510	114	456085	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	375297	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	124633	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.944	65	223328	56.59	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 113.18%		
35) Dibromofluoromethane	7.504	113	169931	54.67	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 109.34%		
50) Toluene-d8	10.025	98	558179	47.26	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 94.52%		
62) 4-Bromofluorobenzene	12.345	95	132821	34.68	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 69.36%#		
Target Compounds						
						Qvalue
16) Acetone	3.744	43	25666	15.52	ug/l	92
18) Methyl Acetate	4.246	43	6369	1.54	ug/l #	89
20) Methylene Chloride	4.455	84	17398	4.90	ug/l #	96
43) Isopropyl Acetate	8.083	43	25252	2.97	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040821\
Data File : VN066547.D
Acq On : 9 Apr 2021 1:19
Operator : JC/MD
Sample : M1961-20
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR2-C-BOTTOM

Quant Time: Apr 09 04:36:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 06 15:38:34 2021
Response via : Initial Calibration

