

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040821\
 Data File : VN066530.D
 Acq On : 8 Apr 2021 10:34
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 09 04:30:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 06 15:38:34 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	50.000	53.732	-7.5	107	0.00
3 P	Chloromethane	50.000	48.208	3.6	105	0.00
4 C	Vinyl Chloride	50.000	47.801	4.4#	99	0.00
5 T	Bromomethane	50.000	45.847	8.3	99	0.00
6 T	Chloroethane	50.000	48.116	3.8	98	0.00
7 T	Trichlorofluoromethane	50.000	48.258	3.5	98	0.00
8 T	Diethyl Ether	50.000	49.683	0.6	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	51.211	-2.4	103	0.00
10 T	Methyl Iodide	50.000	50.466	-0.9	97	0.00
11 T	Tert butyl alcohol	250.000	222.097	11.2	98	0.00
12 CM	1,1-Dichloroethene	50.000	48.590	2.8#	101	0.00
13 T	Acrolein	250.000	218.843	12.5	96	0.00
14 T	Allyl chloride	50.000	47.899	4.2	109	0.00
15 T	Acrylonitrile	250.000	234.232	6.3	97	0.00
16 T	Acetone	250.000	269.853	-7.9	125	0.00
17 T	Carbon Disulfide	50.000	46.407	7.2	103	0.00
18 T	Methyl Acetate	50.000	45.668	8.7	103	0.00
19 T	Methyl tert-butyl Ether	50.000	50.714	-1.4	103	0.00
20 T	Methylene Chloride	50.000	48.209	3.6	102	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.394	3.2	101	0.00
22 T	Diisopropyl ether	50.000	53.929	-7.9	110	0.00
23 T	Vinyl Acetate	250.000	264.262	-5.7	106	0.00
24 P	1,1-Dichloroethane	50.000	50.957	-1.9	106	0.00
25 T	2-Butanone	250.000	248.834	0.5	104	0.00
26 T	2,2-Dichloropropane	50.000	52.073	-4.1	107	0.00
27 T	cis-1,2-Dichloroethene	50.000	49.978	0.0	103	0.00
28 T	Bromochloromethane	50.000	48.013	4.0	109	0.00
29 T	Tetrahydrofuran	250.000	231.798	7.3	95	0.00
30 C	Chloroform	50.000	50.806	-1.6#	104	0.00
31 T	Cyclohexane	50.000	49.658	0.7	108	0.00
32 T	1,1,1-Trichloroethane	50.000	49.710	0.6	101	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.276	3.4	104	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	103	0.00
35 S	Dibromofluoromethane	50.000	47.737	4.5	101	0.00
36 T	1,1-Dichloropropene	50.000	51.373	-2.7	105	0.00
37 T	Ethyl Acetate	50.000	46.272	7.5	99	0.00
38 T	Carbon Tetrachloride	50.000	50.244	-0.5	100	0.00
39 T	Methylcyclohexane	50.000	52.825	-5.7	106	0.00
40 TM	Benzene	50.000	52.569	-5.1	106	0.00
41 T	Methacrylonitrile	50.000	46.760	6.5	99	0.01
42 TM	1,2-Dichloroethane	50.000	51.504	-3.0	106	0.00
43 T	Isopropyl Acetate	50.000	49.777	0.4	102	0.00
44 TM	Trichloroethene	50.000	47.603	4.8	96	0.00
45 C	1,2-Dichloropropane	50.000	53.289	-6.6#	108	0.00
46 T	Dibromomethane	50.000	51.580	-3.2	104	0.00
47 T	Bromodichloromethane	50.000	50.992	-2.0	102	0.00
48 T	Methyl methacrylate	50.000	50.251	-0.5	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040821\
 Data File : VN066530.D
 Acq On : 8 Apr 2021 10:34
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 09 04:30:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 06 15:38:34 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1073.699	-7.4	105	0.00
50 S	Toluene-d8	50.000	49.613	0.8	104	0.00
51 T	4-Methyl-2-Pentanone	250.000	249.755	0.1	97	0.00
52 CM	Toluene	50.000	53.521	-7.0#	104	0.00
53 T	t-1,3-Dichloropropene	50.000	53.264	-6.5	104	0.00
54 T	cis-1,3-Dichloropropene	50.000	53.018	-6.0	106	0.00
55 T	1,1,2-Trichloroethane	50.000	51.005	-2.0	101	0.00
56 T	Ethyl methacrylate	50.000	45.693	8.6	102	0.00
57 T	1,3-Dichloropropane	50.000	52.732	-5.5	105	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	291.876	-16.8	124	0.00
59 T	2-Hexanone	250.000	222.888	10.8	97	0.00
60 T	Dibromochloromethane	50.000	51.090	-2.2	98	0.00
61 T	1,2-Dibromoethane	50.000	50.921	-1.8	100	0.00
62 S	4-Bromofluorobenzene	50.000	51.851	-3.7	105	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	104	0.00
64 T	Tetrachloroethene	50.000	43.248	13.5	87	0.00
65 PM	Chlorobenzene	50.000	50.103	-0.2	100	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.632	-1.3	101	0.00
67 C	Ethyl Benzene	50.000	53.141	-6.3#	102	0.00
68 T	m/p-Xylenes	100.000	110.081	-10.1	103	0.00
69 T	o-Xylene	50.000	53.070	-6.1	103	0.00
70 T	Styrene	50.000	48.018	4.0	103	0.00
71 P	Bromoform	50.000	50.797	-1.6	98	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	102	0.00
73 T	Isopropylbenzene	50.000	52.756	-5.5	103	0.00
74 T	N-amyl acetate	50.000	42.684	14.6	98	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	46.856	6.3	103	0.00
76 T	1,2,3-Trichloropropane	50.000	49.018	2.0	96	0.00
77 T	Bromobenzene	50.000	49.313	1.4	101	0.00
78 T	n-propylbenzene	50.000	54.346	-8.7	104	0.00
79 T	2-Chlorotoluene	50.000	52.112	-4.2	104	0.00
80 T	1,3,5-Trimethylbenzene	50.000	53.218	-6.4	103	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	48.586	2.8	101	0.00
82 T	4-Chlorotoluene	50.000	54.601	-9.2	105	0.00
83 T	tert-Butylbenzene	50.000	49.504	1.0	104	0.00
84 T	1,2,4-Trimethylbenzene	50.000	53.806	-7.6	104	0.00
85 T	sec-Butylbenzene	50.000	53.628	-7.3	105	0.00
86 T	p-Isopropyltoluene	50.000	55.314	-10.6	106	0.00
87 T	1,3-Dichlorobenzene	50.000	49.906	0.2	104	0.00
88 T	1,4-Dichlorobenzene	50.000	50.164	-0.3	105	0.00
89 T	n-Butylbenzene	50.000	52.524	-5.0	111	0.00
90 T	Hexachloroethane	50.000	50.555	-1.1	109	0.00
91 T	1,2-Dichlorobenzene	50.000	49.348	1.3	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	41.099	17.8	86	0.00
93 T	1,2,4-Trichlorobenzene	50.000	44.886	10.2	99	0.00
94 T	Hexachlorobutadiene	50.000	46.339	7.3	102	0.00
95 T	Naphthalene	50.000	39.991	20.0	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040821\
Data File : VN066530.D
Acq On : 8 Apr 2021 10:34
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Apr 09 04:30:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 06 15:38:34 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	43.295	13.4	96	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6