


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040821\
 Data File : VN066529.D
 Acq On : 8 Apr 2021 9:29
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
 Title : SW846 8260
 Last Update : Tue Apr 06 15:38:34 2021



AutoFind: Scans 3973, 3974, 3975; Background Corrected with Scan 3958

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.2	6130	PASS
75	95	30	60	50.2	15218	PASS
95	95	100	100	100.0	30320	PASS
96	95	5	9	7.6	2305	PASS
173	174	0.00	2	1.1	280	PASS
174	95	50	100	85.9	26053	PASS
175	174	5	9	7.9	2052	PASS
176	174	95	101	97.0	25275	PASS
177	176	5	9	6.9	1736	PASS