

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040822\
 Data File : VN071908.D
 Acq On : 08 Apr 2022 11:20
 Operator : JC\MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Apr 09 05:19:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032422W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Mar 24 12:03:44 2022
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	77	0.00
2 M	Dichlorodifluoromethane	1.867	1.326	29.0	53	0.00
3 M	Chloromethane	2.201	1.608	26.9	57	0.00
4 M	Vinyl Chloride	2.524	1.702	32.6#	53	0.00
5 M	Bromomethane	1.487	1.228	17.4	57	-0.01
6 M	Chloroethane	1.616	1.163	28.0	57	0.00
7 M	Trichlorofluoromethane	2.943	2.418	17.8	64	-0.01
8 T	Diethyl Ether	1.236	1.136	8.1	72	0.00
9	1,1,2-Trichlorotrifluoroeth	1.762	1.599	9.3	70	0.00
10 M	1,1-Dichloroethene	1.738	1.487	14.4	67	0.00
11	Methyl Iodide	2.547	2.071	18.7	65	0.00
12	Methyl Acetate	2.449	2.393	2.3	77	0.00
13 M	Acrolein	0.261	0.284	-8.8	90	0.00
14 M	Acrylonitrile	1.109	1.095	1.3	79	0.00
15 M	Acetone	0.319	0.271	15.0	64	0.00
16 M	Carbon Disulfide	4.525	2.771	38.8#	49	0.00
17	Allyl chloride	2.819	2.382	15.5	68	0.00
18 M	Methylene Chloride	2.085	1.875	10.1	71	0.00
19 M	trans-1,2-Dichloroethene	1.859	1.592	14.4	68	0.00
20 T	Diisopropyl ether	5.898	5.553	5.8	73	0.00
21 M	1,1-Dichloroethane	3.561	3.250	8.7	72	0.00
22 M	cis-1,2-Dichloroethene	2.258	2.086	7.6	73	0.00
23 M	tert-Butyl Alcohol	0.396	0.357	9.8	74	0.00
24 M	Methyl tert-Butyl Ether	6.273	6.055	3.5	76	0.00
25 M	Chloroform	3.582	3.445	3.8	75	0.00
26	Cyclohexane	3.078	2.464	19.9	63	0.00
27 s	1,2-Dichloroethane-d4	2.114	2.020	4.4	73	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	79	0.00
29	1,1-Dichloropropene	0.486	0.404	16.9	67	0.00
30 M	2-Butanone	0.270	0.240	11.1	71	0.00
31	2,2-Dichloropropane	0.553	0.463	16.3	68	0.00
32 M	1,1,1-Trichloroethane	0.566	0.502	11.3	72	0.00
33 M	Carbon Tetrachloride	0.479	0.416	13.2	69	0.00
34 M	Benzene	1.509	1.354	10.3	71	0.00
35	Methacrylonitrile	0.259	0.256	1.2	76	0.00
36 M	1,2-Dichloroethane	0.526	0.465	11.6	71	0.00
37 M	Trichloroethene	0.389	0.344	11.6	72	0.00
38	Methylcyclohexane	0.612	0.479	21.7	63	0.00
39 M	1,2-Dichloropropane	0.377	0.357	5.3	75	0.00
40	Dibromomethane	0.255	0.233	8.6	73	0.00
41 M	Bromodichloromethane	0.511	0.472	7.6	75	0.00
42 M	Vinyl Acetate	0.926	0.818	11.7	72	0.00
43	Ethyl Acetate	0.524	0.475	9.4	72	0.00
44	Isopropyl Acetate	0.841	0.739	12.1	73	0.00
45 T	1,4-Dioxane	0.008	0.008#	0.0	76	0.00
46	Methyl methacrylate	0.369	0.316	14.4	73	0.00
47	n-amyl Acetate	0.561	0.466	16.9	74	0.00
48 M	t-1,3-Dichloropropene	0.556	0.498	10.4	76	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040822\
 Data File : VN071908.D
 Acq On : 08 Apr 2022 11:20
 Operator : JC\MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Apr 09 05:19:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032422W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Mar 24 12:03:44 2022
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	cis-1,3-Dichloropropene	0.608	0.546	10.2	74	0.00
50 M	1,1,2-Trichloroethane	0.382	0.383	-0.3	80	0.00
51	Ethyl methacrylate	0.590	0.547	7.3	79	0.00
52	1,3-Dichloropropane	0.652	0.623	4.4	76	0.00
53 M	Dibromochloromethane	0.395	0.384	2.8	80	0.00
54 M	1,2-Dibromoethane	0.393	0.365	7.1	76	0.00
55 M	2-Chloroethyl vinyl ether	0.159	0.130	18.2	68	0.00
56 M	Bromoform	0.305	0.289	5.2	80	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	81	0.00
58 M	4-Methyl-2-Pentanone	0.569	0.528	7.2	76	0.00
59 M	2-Hexanone	0.416	0.374	10.1	74	0.00
60 S	4-Bromofluorobenzene	0.445	0.439	1.3	81	0.00
61 M	Tetrachloroethene	0.365	0.340	6.8	75	0.00
62 M	Toluene	1.819	1.674	8.0	74	0.00
63 S	Toluene-d8	1.362	1.357	0.4	78	0.00
64 M	Chlorobenzene	1.122	1.055	6.0	78	0.00
65	1,1,1,2-Tetrachloroethane	0.412	0.394	4.4	78	0.00
66 M	Ethyl Benzene	1.977	1.785	9.7	75	0.00
67 M	m/p-Xylenes	0.769	0.689	10.4	73	0.00
68 M	o-Xylene	0.764	0.690	9.7	75	0.00
69 M	Styrene	1.203	1.083	10.0	77	0.00
70	Isopropylbenzene	1.918	1.783	7.0	77	0.00
71 M	1,1,2,2-Tetrachloroethane	0.624	0.618	1.0	79	0.00
72	1,2,3-Trichloropropane	0.536	0.528	1.5	78	0.00
73	Bromobenzene	0.468	0.431	7.9	78	0.00
74	n-propylbenzene	2.094	1.857	11.3	76	0.00
75	2-Chlorotoluene	1.308	1.194	8.7	76	0.00
76	1,3,5-Trimethylbenzene	1.539	1.421	7.7	76	0.00
77	t-1,4-Dichloro-2-butene	0.188	0.153	18.6	75	0.00
78	4-Chlorotoluene	1.222	1.091	10.7	76	0.00
79	tert-butylbenzene	1.389	1.311	5.6	78	0.00
80	1,2,4-Trimethylbenzene	1.493	1.378	7.7	77	0.00
81	sec-Butylbenzene	1.866	1.707	8.5	78	0.00
82	p-Isopropyltoluene	1.516	1.367	9.8	78	0.00
83 M	1,3-Dichlorobenzene	0.766	0.698	8.9	79	0.00
84 M	1,4-Dichlorobenzene	0.731	0.651	10.9	77	0.00
85	n-Butylbenzene	1.126	0.929	17.5	76	0.00
86 T	Hexachloroethane	0.283	0.255	9.9	77	0.00
87 M	1,2-Dichlorobenzene	0.740	0.704	4.9	78	0.00
88	1,2-Dibromo-3-Chloropropane	0.111	0.097	12.6	76	0.00
89	1,2,4-Trichlorobenzene	0.342	0.322	5.8	88	0.00
90	Hexachlorobutadiene	0.203	0.203	0.0	84	0.00
91 M	Naphthalene	0.951	0.908	4.5	97	0.00
92	1,2,3-Trichlorobenzene	0.336	0.346	-3.0	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0