

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040824\
 Data File : VN081675.D
 Acq On : 08 Apr 2024 15:36
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Apr 09 01:56:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	395274	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.100	114	715965	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	649388	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	261154	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	255199	50.627	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.260%
35) Dibromofluoromethane	8.165	113	221012	48.913	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.820%
50) Toluene-d8	10.565	98	846608	51.832	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.660%
62) 4-Bromofluorobenzene	12.853	95	278575	48.403	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	96.800%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	18137	11.133	ug/l	99
52) Toluene	10.629	92	17468	1.236	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040824\
Data File : VN081675.D
Acq On : 08 Apr 2024 15:36
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Apr 09 01:56:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

