

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040824\
 Data File : VN081676.D
 Acq On : 08 Apr 2024 16:13
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 09 01:56:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	81	0.00
2 T	Dichlorodifluoromethane	0.673	0.714	-6.1	90	0.00
3 P	Chloromethane	0.670	0.674	-0.6	88	0.00
4 C	Vinyl Chloride	0.752	0.688	8.5#	79	0.00
5 T	Bromomethane	0.532	0.480	9.8	78	0.00
6 T	Chloroethane	0.517	0.482	6.8	80	0.00
7 T	Trichlorofluoromethane	1.121	1.243	-10.9	93	0.00
8 T	Diethyl Ether	0.371	0.399	-7.5	87	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.637	0.709	-11.3	95	0.00
10 T	Methyl Iodide	0.634	0.822	-29.7#	99	0.00
11 T	Tert butyl alcohol	0.111	0.111	0.0	80	0.00
12 CM	1,1-Dichloroethene	0.590	0.658	-11.5#	93	0.00
13 T	Acrolein	0.100	0.083	17.0	62	0.00
14 T	Allyl chloride	0.739	0.802	-8.5	89	0.00
15 T	Acrylonitrile	0.297	0.271	8.8	76	0.00
16 T	Acetone	0.206	0.182	11.7	77	0.00
17 T	Carbon Disulfide	1.693	1.816	-7.3	93	0.00
18 T	Methyl Acetate	0.635	0.534	15.9	68	0.00
19 T	Methyl tert-butyl Ether	1.847	2.033	-10.1	88	0.00
20 T	Methylene Chloride	0.675	0.739	-9.5	96	0.00
21 T	trans-1,2-Dichloroethene	0.658	0.708	-7.6	93	0.00
22 T	Diisopropyl ether	1.718	1.889	-10.0	86	0.00
23 T	Vinyl Acetate	1.368	1.463	-6.9	82	0.00
24 P	1,1-Dichloroethane	1.151	1.213	-5.4	89	0.00
25 T	2-Butanone	0.353	0.306	13.3	70	0.00
26 T	2,2-Dichloropropane	0.998	1.161	-16.3	92	0.00
27 T	cis-1,2-Dichloroethene	0.744	0.870	-16.9	97	0.00
28 T	Bromochloromethane	0.469	0.418	10.9	74	0.00
29 T	Tetrahydrofuran	0.238	0.216	9.2	72	0.00
30 C	Chloroform	1.232	1.367	-11.0#	92	0.00
31 T	Cyclohexane	1.038	1.078	-3.9	91	0.00
32 T	1,1,1-Trichloroethane	1.118	1.251	-11.9	93	0.00
33 S	1,2-Dichloroethane-d4	0.638	0.592	7.2	74	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	87	0.00
35 S	Dibromofluoromethane	0.316	0.307	2.8	79	0.00
36 T	1,1-Dichloropropene	0.531	0.530	0.2	85	0.00
37 T	Ethyl Acetate	0.455	0.396	13.0	74	0.00
38 T	Carbon Tetrachloride	0.582	0.640	-10.0	95	0.00
39 T	Methylcyclohexane	0.593	0.689	-16.2	94	0.00
40 TM	Benzene	1.553	1.665	-7.2	91	0.00
41 T	Methacrylonitrile	0.240	0.206	14.2	76	0.00
42 TM	1,2-Dichloroethane	0.546	0.518	5.1	83	0.00
43 T	Isopropyl Acetate	0.754	0.778	-3.2	86	0.00
44 TM	Trichloroethene	0.437	0.439	-0.5	88	0.00
45 C	1,2-Dichloropropane	0.376	0.373	0.8#	84	0.00
46 T	Dibromomethane	0.286	0.294	-2.8	91	0.00
47 T	Bromodichloromethane	0.559	0.609	-8.9	92	0.00
48 T	Methyl methacrylate	0.348	0.317	8.9	75	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040824\
 Data File : VN081676.D
 Acq On : 08 Apr 2024 16:13
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 09 01:56:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.007	12.5	70	0.00
50 S	Toluene-d8	1.141	1.074	5.9	76	0.00
51 T	4-Methyl-2-Pentanone	0.460	0.408	11.3	73	0.00
52 CM	Toluene	0.987	1.104	-11.9#	95	0.00
53 T	t-1,3-Dichloropropene	0.603	0.623	-3.3	85	0.00
54 T	cis-1,3-Dichloropropene	0.647	0.662	-2.3	86	0.00
55 T	1,1,2-Trichloroethane	0.397	0.398	-0.3	88	0.00
56 T	Ethyl methacrylate	0.529	0.590	-11.5	87	0.00
57 T	1,3-Dichloropropane	0.651	0.658	-1.1	85	0.00
58 T	2-Chloroethyl Vinyl ether	0.250	0.271	-8.4	76	0.00
59 T	2-Hexanone	0.340	0.298	12.4	70	0.00
60 T	Dibromochloromethane	0.437	0.489	-11.9	96	0.00
61 T	1,2-Dibromoethane	0.405	0.408	-0.7	84	0.00
62 S	4-Bromofluorobenzene	0.402	0.412	-2.5	81	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	85	0.00
64 T	Tetrachloroethene	0.478	0.541	-13.2	99	0.00
65 PM	Chlorobenzene	1.182	1.340	-13.4	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.438	0.469	-7.1	89	0.00
67 C	Ethyl Benzene	2.053	2.333	-13.6#	92	0.00
68 T	m/p-Xylenes	0.770	0.907	-17.8	94	0.00
69 T	o-Xylene	0.756	0.870	-15.1	94	0.00
70 T	Styrene	1.219	1.497	-22.8	96	0.00
71 P	Bromoform	0.329	0.340	-3.3	85	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
73 T	Isopropylbenzene	4.262	4.658	-9.3	91	0.00
74 T	N-amyl acetate	1.542	1.484	3.8	78	0.00
75 P	1,1,2,2-Tetrachloroethane	1.253	1.035	17.4	75	0.00
76 T	1,2,3-Trichloropropane	1.196	0.998	16.6	76	0.00
77 T	Bromobenzene	1.053	1.108	-5.2	94	0.00
78 T	n-propylbenzene	5.033	5.456	-8.4	89	0.00
79 T	2-Chlorotoluene	3.012	3.191	-5.9	91	0.00
80 T	1,3,5-Trimethylbenzene	3.520	3.890	-10.5	90	0.00
81 T	trans-1,4-Dichloro-2-butene	0.443	0.437	1.4	81	0.00
82 T	4-Chlorotoluene	2.970	3.187	-7.3	90	0.00
83 T	tert-Butylbenzene	3.036	3.411	-12.4	94	0.00
84 T	1,2,4-Trimethylbenzene	3.538	3.979	-12.5	93	0.00
85 T	sec-Butylbenzene	4.340	4.817	-11.0	92	0.00
86 T	p-Isopropyltoluene	3.547	4.129	-16.4	95	0.00
87 T	1,3-Dichlorobenzene	1.934	2.084	-7.8	97	0.00
88 T	1,4-Dichlorobenzene	2.003	2.065	-3.1	95	0.00
89 T	n-Butylbenzene	3.146	3.521	-11.9	90	0.00
90 T	Hexachloroethane	0.672	0.732	-8.9	95	0.00
91 T	1,2-Dichlorobenzene	1.862	1.944	-4.4	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.267	0.219	18.0	72	0.00
93 T	1,2,4-Trichlorobenzene	1.077	1.148	-6.6	91	0.00
94 T	Hexachlorobutadiene	0.472	0.446	5.5	84	0.00
95 T	Naphthalene	3.488	3.582	-2.7	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040824\
Data File : VN081676.D
Acq On : 08 Apr 2024 16:13
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Apr 09 01:56:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.083	1.136	-4.9	91	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6