

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040921\
 Data File : VN066557.D
 Acq On : 9 Apr 2021 12:50
 Operator : JC/MD
 Sample : M1990-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ELEVATOR-PIT-STANDING-WATER

Manual Integrations
 APPROVED

MMDadoda
 4/12/2021 11:44:16 AM

Quant Time: Apr 09 14:01:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032421W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Mar 25 05:52:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.104	128	47892	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.512	114	228927	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.350	117	203582	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.949	65	114834	31.20	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	104.00%
60) 4-Bromofluorobenzene	12.348	95	65440	20.60	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	68.67%
63) Toluene-d8	10.028	98	293633	30.73	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	102.43%
Target Compounds						Qvalue
15) Acetone	3.741	58	4447m	9.64	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040921\
Data File : VN066557.D
Acq On : 9 Apr 2021 12:50
Operator : JC/MD
Sample : M1990-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ELEVATOR-PIT-STANDING-WATER

Manual Integrations
APPROVED

MMDadoda
4/12/2021 11:44:16 AM

Quant Time: Apr 09 14:01:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032421W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Mar 25 05:52:50 2021
Response via : Initial Calibration

