

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040921\
 Data File : VN066558.D
 Acq On : 9 Apr 2021 13:12
 Operator : JC/MD
 Sample : M1914-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 001-WILLETS-PT-BLVD(APR)

Quant Time: Apr 09 14:01:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032421W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Mar 25 05:52:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.107	128	47926	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.515	114	222487	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.350	117	196153	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.949	65	111670	30.31	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	101.03%
60) 4-Bromofluorobenzene	12.348	95	64482	21.06	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	70.20%
63) Toluene-d8	10.028	98	287269	31.21	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	104.03%
Target Compounds						
					Qvalue	
15) Acetone	3.736	58	127267	275.82	ug/l	97
30) 2-Butanone	6.745	43	9341	4.79	ug/l #	64
58) 4-Methyl-2-Pentanone	9.923	43	19294	4.91	ug/l #	88
62) Toluene	10.092	91	4867	0.45	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040921\
Data File : VN066558.D
Acq On : 9 Apr 2021 13:12
Operator : JC/MD
Sample : M1914-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
001-WILLETS-PT-BLVD(APR)

Quant Time: Apr 09 14:01:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032421W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Mar 25 05:52:50 2021
Response via : Initial Calibration

