

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041023\
 Data File : VN077287.D
 Acq On : 10 Apr 2023 11:35
 Operator : JC\MD
 Sample : VN0410WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0410WBL01

Quant Time: Apr 11 01:22:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N040723W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Apr 07 13:04:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.819	128	56079	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.101	114	416164	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.866	117	415101	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.583	65	139199	29.268	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	97.567%
60) 4-Bromofluorobenzene	12.848	95	198334	27.632	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	92.100%
63) Toluene-d8	10.566	98	449787	30.823	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	102.733%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041023\
Data File : VN077287.D
Acq On : 10 Apr 2023 11:35
Operator : JC\MD
Sample : VN0410WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0410WBL01

Quant Time: Apr 11 01:22:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N040723W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Apr 07 13:04:49 2023
Response via : Initial Calibration

