

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041122\
 Data File : VN071928.D
 Acq On : 11 Apr 2022 12:04
 Operator : JC\MD
 Sample : VN0411MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0411MBL01

Quant Time: Apr 11 12:28:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	531100	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	888848	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	909367	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	324325	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	286374	47.232	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.460%
35) Dibromofluoromethane	8.016	113	255128	48.778	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.560%
50) Toluene-d8	10.439	98	1109028	51.833	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.660%
62) 4-Bromofluorobenzene	12.727	95	365535	52.824	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	105.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041122\
Data File : VN071928.D
Acq On : 11 Apr 2022 12:04
Operator : JC\MD
Sample : VN0411MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0411MBL01

Quant Time: Apr 11 12:28:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

