

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041122\
 Data File : VN071929.D
 Acq On : 11 Apr 2022 12:28
 Operator : JC\MD
 Sample : VN0411WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0411WBL01

Quant Time: Apr 11 13:03:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	645589	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1102283	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1157299	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	434328	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	351274	47.662	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.320%
35) Dibromofluoromethane	8.016	113	316243	48.755	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.520%
50) Toluene-d8	10.439	98	1385185	52.204	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.400%
62) 4-Bromofluorobenzene	12.727	95	484933	56.509	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	113.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041122\
Data File : VN071929.D
Acq On : 11 Apr 2022 12:28
Operator : JC\MD
Sample : VN0411WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0411WBL01

Quant Time: Apr 11 13:03:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

