

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041122\
 Data File : VN071935.D
 Acq On : 11 Apr 2022 14:52
 Operator : JC\MD
 Sample : N2335-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 A350-CU

Quant Time: Apr 12 08:39:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	661085	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1130710	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1202124	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	473562	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	357905	47.423	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.840%
35) Dibromofluoromethane	8.022	113	319286	47.987	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.980%
50) Toluene-d8	10.439	98	1428148	52.470	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.940%
62) 4-Bromofluorobenzene	12.727	95	517631	58.803	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	117.600%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	46467	14.407	ug/l	99
20) Methylene Chloride	5.098	84	23899	3.515	ug/l #	84
43) Isopropyl Acetate	8.557	43	91892	5.432	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041122\
Data File : VN071935.D
Acq On : 11 Apr 2022 14:52
Operator : JC\MD
Sample : N2335-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
A350-CU

Quant Time: Apr 12 08:39:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

