

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041122\
 Data File : VN071937.D
 Acq On : 11 Apr 2022 15:40
 Operator : JC\MD
 Sample : N2337-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 B250-CU

Quant Time: Apr 12 08:40:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	744207	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1268792	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1319861	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	503523	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	405104	47.682	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.360%
35) Dibromofluoromethane	8.016	113	355919	47.671	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.340%
50) Toluene-d8	10.439	98	1611498	52.763	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.520%
62) 4-Bromofluorobenzene	12.727	95	557329	56.422	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	112.840%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	41536	11.440	ug/l	92
20) Methylene Chloride	5.093	84	18992	2.482	ug/l	97
43) Isopropyl Acetate	8.569	43	170426	8.979	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041122\
Data File : VN071937.D
Acq On : 11 Apr 2022 15:40
Operator : JC\MD
Sample : N2337-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
B250-CU

Quant Time: Apr 12 08:40:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

