

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041122\
 Data File : VN071940.D
 Acq On : 11 Apr 2022 16:52
 Operator : JC\MD
 Sample : PB143992ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143992ZHE#02

Manual Integrations
 APPROVED

Reviewed By : John Carlone 04/12/2022
 Supervised By : Mahesh Dadoda 04/12/2022

Quant Time: Apr 12 08:41:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	713710	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1219793	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1280953	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	488888	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	390297	47.902	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	95.800%		
35) Dibromofluoromethane	8.016	113	344952	48.058	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	96.120%		
50) Toluene-d8	10.439	98	1547820	52.714	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	105.420%		
62) 4-Bromofluorobenzene	12.727	95	540982	56.967	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	113.940%		
Target Compounds						
16) Acetone	4.292	43	18052	5.184	ug/l	98
43) Isopropyl Acetate	8.563	43	170198m	9.327	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041122\
Data File : VN071940.D
Acq On : 11 Apr 2022 16:52
Operator : JC\MD
Sample : PB143992ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143992ZHE#02

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/12/2022
Supervised By :Mahesh Dadoda 04/12/2022

Quant Time: Apr 12 08:41:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

