

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041122\
 Data File : VN071950.D
 Acq On : 11 Apr 2022 20:53
 Operator : JC\MD
 Sample : PB143992ZHE#12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143992ZHE#12

Quant Time: Apr 12 08:43:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	620813	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1087430	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1146229	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	440138	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	339860	47.953	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.900%
35) Dibromofluoromethane	8.022	113	305014	47.667	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.340%
50) Toluene-d8	10.439	98	1379861	52.714	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.420%
62) 4-Bromofluorobenzene	12.727	95	486891	57.512	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	115.020%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	17188	5.675	ug/l	94
18) Methyl Acetate	4.863	43	28022	2.292	ug/l	99
20) Methylene Chloride	5.104	84	7111	1.114	ug/l #	87
43) Isopropyl Acetate	8.569	43	214072	13.159	ug/l #	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041122\
Data File : VN071950.D
Acq On : 11 Apr 2022 20:53
Operator : JC\MD
Sample : PB143992ZHE#12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143992ZHE#12

Quant Time: Apr 12 08:43:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

