

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041122\
 Data File : VN071951.D
 Acq On : 11 Apr 2022 21:18
 Operator : JC\MD
 Sample : PB143992ZHE#13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143992ZHE#13

Quant Time: Apr 12 08:43:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	701911	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1216790	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1286550	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	503072	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	383995	47.921	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.840%
35) Dibromofluoromethane	8.016	113	345113	48.199	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.400%
50) Toluene-d8	10.439	98	1545720	52.772	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.540%
62) 4-Bromofluorobenzene	12.727	95	555474	58.638	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	117.280%
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	24354	7.112	ug/l	97
18) Methyl Acetate	4.857	43	27232	1.750	ug/l	100
20) Methylene Chloride	5.104	84	15445	2.140	ug/l	86
43) Isopropyl Acetate	8.569	43	211592	11.624	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041122\
Data File : VN071951.D
Acq On : 11 Apr 2022 21:18
Operator : JC\MD
Sample : PB143992ZHE#13
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143992ZHE#13

Quant Time: Apr 12 08:43:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

