

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041123\
 Data File : VN077296.D
 Acq On : 11 Apr 2023 11:47
 Operator : JC\MD
 Sample : PB151999TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB151999TB

Quant Time: Apr 12 04:38:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 06 08:27:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	468228	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.101	114	923895	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	863521	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	308150	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	303178	47.376	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	94.760%		
35) Dibromofluoromethane	8.172	113	279375	47.236	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.480%		
50) Toluene-d8	10.572	98	1151397	48.358	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	96.720%		
62) 4-Bromofluorobenzene	12.854	95	395917	49.997	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	100.000%		
Target Compounds						
					Qvalue	
16) Acetone	4.443	43	54314	23.019	ug/l	97
20) Methylene Chloride	5.284	84	56019	8.407	ug/l	97
37) Ethyl Acetate	7.566	43	96943	12.386	ug/l	97
43) Isopropyl Acetate	8.689	43	450688	31.207	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041123\
Data File : VN077296.D
Acq On : 11 Apr 2023 11:47
Operator : JC\MD
Sample : PB151999TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB151999TB

Quant Time: Apr 12 04:38:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 06 08:27:03 2023
Response via : Initial Calibration

