

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041123\
 Data File : VN077298.D
 Acq On : 11 Apr 2023 12:35
 Operator : JC\MD
 Sample : PB151999ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB151999ZHE#02

Quant Time: Apr 12 04:39:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 06 08:27:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	489286	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	928808	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	881473	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	307176	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	305376	45.665	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.340%
35) Dibromofluoromethane	8.172	113	281592	47.359	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.720%
50) Toluene-d8	10.572	98	1172078	48.966	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.940%
62) 4-Bromofluorobenzene	12.848	95	362913	45.587	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.180%
Target Compounds						
					Qvalue	
16) Acetone	4.443	43	56726	23.007	ug/l	89
20) Methylene Chloride	5.290	84	54578	7.838	ug/l #	93
37) Ethyl Acetate	7.566	43	90603	11.515	ug/l	98
43) Isopropyl Acetate	8.689	43	457080	31.482	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041123\
Data File : VN077298.D
Acq On : 11 Apr 2023 12:35
Operator : JC\MD
Sample : PB151999ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB151999ZHE#02

Quant Time: Apr 12 04:39:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 06 08:27:03 2023
Response via : Initial Calibration

