

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041123\
 Data File : VN077299.D
 Acq On : 11 Apr 2023 12:59
 Operator : JC\MD
 Sample : PB151999ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB151999ZHE#03

Quant Time: Apr 12 04:40:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 06 08:27:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	510475	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	968819	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	910965	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	321056	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	316780	45.404	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.800%
35) Dibromofluoromethane	8.172	113	293706	47.356	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.720%
50) Toluene-d8	10.572	98	1220467	48.882	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.760%
62) 4-Bromofluorobenzene	12.854	95	379403	45.690	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.380%
Target Compounds						
					Qvalue	
16) Acetone	4.437	43	51559	20.043	ug/l	100
20) Methylene Chloride	5.284	84	56505	7.778	ug/l	96
37) Ethyl Acetate	7.566	43	88839	10.825	ug/l	98
43) Isopropyl Acetate	8.695	43	464857	30.696	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041123\
Data File : VN077299.D
Acq On : 11 Apr 2023 12:59
Operator : JC\MD
Sample : PB151999ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB151999ZHE#03

Quant Time: Apr 12 04:40:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 06 08:27:03 2023
Response via : Initial Calibration

