

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041123\
 Data File : VN077301.D
 Acq On : 11 Apr 2023 13:47
 Operator : JC\MD
 Sample : 02262-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JPP-1-040723

Quant Time: Apr 12 04:41:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 06 08:27:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	489472	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	935170	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	884307	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	312645	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	311516	46.566	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.140%
35) Dibromofluoromethane	8.172	113	278931	46.592	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.180%
50) Toluene-d8	10.566	98	1193602	49.526	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.060%
62) 4-Bromofluorobenzene	12.854	95	392134	48.922	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.840%
Target Compounds						
					Qvalue	
8) Diethyl Ether	3.972	74	33162	8.695	ug/l	98
16) Acetone	4.449	43	43015	17.439	ug/l	91
20) Methylene Chloride	5.284	84	40877	5.868	ug/l	93
37) Ethyl Acetate	7.566	43	75224	9.496	ug/l	97
43) Isopropyl Acetate	8.689	43	348552	23.844	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041123\
Data File : VN077301.D
Acq On : 11 Apr 2023 13:47
Operator : JC\MD
Sample : 02262-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JPP-1-040723

Quant Time: Apr 12 04:41:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 06 08:27:03 2023
Response via : Initial Calibration

