

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041123\
 Data File : VN077291.D
 Acq On : 11 Apr 2023 09:16
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 12 04:34:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 06 08:27:03 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.638	0.543	14.9	87	0.00
3 P	Chloromethane	0.791	0.627	20.7	87	0.00
4 C	Vinyl Chloride	1.072	0.914	14.7#	86	0.00
5 T	Bromomethane	0.777	0.627	19.3	88	0.00
6 T	Chloroethane	0.765	0.619	19.1	87	0.00
7 T	Trichlorofluoromethane	0.930	0.872	6.2	96	0.00
8 T	Diethyl Ether	0.390	0.355	9.0	93	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.569	0.527	7.4	99	0.00
10 T	Methyl Iodide	0.692	0.647	6.5	90	0.00
11 T	Tert butyl alcohol	0.139	0.104	25.2#	74	-0.01
12 CM	1,1-Dichloroethene	0.586	0.530	9.6#	96	0.00
13 T	Acrolein	0.132	0.104	21.2	87	0.00
14 T	Allyl chloride	0.885	0.708	20.0	87	0.00
15 T	Acrylonitrile	0.308	0.252	18.2	80	0.00
16 T	Acetone	0.252	0.246	2.4	102	0.00
17 T	Carbon Disulfide	1.579	1.435	9.1	92	0.00
18 T	Methyl Acetate	0.960	0.754	21.5	78	0.00
19 T	Methyl tert-butyl Ether	2.078	1.908	8.2	91	0.00
20 T	Methylene Chloride	0.712	0.609	14.5	92	0.00
21 T	trans-1,2-Dichloroethene	0.641	0.596	7.0	94	0.00
22 T	Diisopropyl ether	1.774	1.624	8.5	90	0.00
23 T	Vinyl Acetate	1.199	1.072	10.6	84	0.00
24 P	1,1-Dichloroethane	1.145	1.045	8.7	91	0.00
25 T	2-Butanone	0.409	0.347	15.2	86	0.00
26 T	2,2-Dichloropropane	1.050	1.031	1.8	99	0.00
27 T	cis-1,2-Dichloroethene	0.783	0.734	6.3	96	0.00
28 T	Bromochloromethane	0.461	0.454	1.5	115	0.00
29 T	Tetrahydrofuran	0.263	0.207	21.3	77	0.00
30 C	Chloroform	1.221	1.140	6.6#	94	0.00
31 T	Cyclohexane	1.062	0.928	12.6	95	0.00
32 T	1,1,1-Trichloroethane	1.111	1.044	6.0	96	0.00
33 S	1,2-Dichloroethane-d4	0.683	0.642	6.0	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	99	0.00
35 S	Dibromofluoromethane	0.320	0.303	5.3	98	0.00
36 T	1,1-Dichloropropene	0.473	0.459	3.0	95	0.00
37 T	Ethyl Acetate	0.424	0.351	17.2	81	0.00
38 T	Carbon Tetrachloride	0.485	0.471	2.9	96	0.00
39 T	Methylcyclohexane	0.638	0.624	2.2	98	0.00
40 TM	Benzene	1.477	1.386	6.2	92	0.00
41 T	Methacrylonitrile	0.228	0.184	19.3	79	0.00
42 TM	1,2-Dichloroethane	0.454	0.428	5.7	92	0.00
43 T	Isopropyl Acetate	0.782	0.648	17.1	83	0.00
44 TM	Trichloroethene	0.353	0.345	2.3	96	0.00
45 C	1,2-Dichloropropane	0.351	0.340	3.1#	95	0.00
46 T	Dibromomethane	0.264	0.244	7.6	92	0.00
47 T	Bromodichloromethane	0.528	0.495	6.3	95	0.00
48 T	Methyl methacrylate	0.333	0.291	12.6	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041123\
 Data File : VN077291.D
 Acq On : 11 Apr 2023 09:16
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleID :
 VSTDCCC050

Quant Time: Apr 12 04:34:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 06 08:27:03 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.006	25.0	80	0.00
50 S	Toluene-d8	1.289	1.248	3.2	98	0.00
51 T	4-Methyl-2-Pentanone	0.452	0.378	16.4	79	0.00
52 CM	Toluene	0.967	0.945	2.3#	94	0.00
53 T	t-1,3-Dichloropropene	0.573	0.567	1.0	93	0.00
54 T	cis-1,3-Dichloropropene	0.611	0.597	2.3	94	0.00
55 T	1,1,2-Trichloroethane	0.379	0.350	7.7	91	0.00
56 T	Ethyl methacrylate	0.591	0.549	7.1	88	0.00
57 T	1,3-Dichloropropane	0.627	0.589	6.1	91	0.00
58 T	2-Chloroethyl Vinyl ether	0.176	0.164	6.8	86	0.00
59 T	2-Hexanone	0.337	0.288	14.5	82	0.00
60 T	Dibromochloromethane	0.391	0.394	-0.8	94	0.00
61 T	1,2-Dibromoethane	0.379	0.365	3.7	92	0.00
62 S	4-Bromofluorobenzene	0.429	0.440	-2.6	102	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	100	0.00
64 T	Tetrachloroethene	0.309	0.307	0.6	100	0.00
65 PM	Chlorobenzene	1.101	1.070	2.8	95	0.00
66 T	1,1,1,2-Tetrachloroethane	0.417	0.384	7.9	93	0.00
67 C	Ethyl Benzene	2.031	1.965	3.2#	95	0.00
68 T	m/p-Xylenes	0.798	0.777	2.6	94	0.00
69 T	o-Xylene	0.787	0.767	2.5	94	0.00
70 T	Styrene	1.267	1.244	1.8	92	0.00
71 P	Bromoform	0.293	0.279	4.8	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	4.856	4.458	8.2	95	0.00
74 T	N-amyl acetate	1.840	1.443	21.6	83	0.00
75 P	1,1,2,2-Tetrachloroethane	1.524	1.262	17.2	88	0.00
76 T	1,2,3-Trichloropropane	1.191	0.907	23.8	73	0.00
77 T	Bromobenzene	0.987	0.901	8.7	95	0.00
78 T	n-propylbenzene	5.366	5.136	4.3	96	0.00
79 T	2-Chlorotoluene	3.231	2.952	8.6	95	0.00
80 T	1,3,5-Trimethylbenzene	3.879	3.725	4.0	94	0.00
81 T	trans-1,4-Dichloro-2-butene	0.529	0.496	6.2	96	0.00
82 T	4-Chlorotoluene	3.046	2.892	5.1	95	0.00
83 T	tert-Butylbenzene	3.453	3.136	9.2	94	0.00
84 T	1,2,4-Trimethylbenzene	3.832	3.668	4.3	93	0.00
85 T	sec-Butylbenzene	4.812	4.666	3.0	97	0.00
86 T	p-Isopropyltoluene	3.782	3.814	-0.8	96	0.00
87 T	1,3-Dichlorobenzene	1.827	1.742	4.7	94	0.00
88 T	1,4-Dichlorobenzene	1.802	1.725	4.3	94	0.00
89 T	n-Butylbenzene	3.139	3.323	-5.9	99	0.00
90 T	Hexachloroethane	0.827	0.765	7.5	96	0.00
91 T	1,2-Dichlorobenzene	1.828	1.716	6.1	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.289	0.224	22.5	78	0.00
93 T	1,2,4-Trichlorobenzene	0.660	0.745	-12.9	91	0.00
94 T	Hexachlorobutadiene	0.382	0.366	4.2	101	0.00
95 T	Naphthalene	2.478	2.590	-4.5	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041123\
Data File : VN077291.D
Acq On : 11 Apr 2023 09:16
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Apr 12 04:34:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 06 08:27:03 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.631	0.686	-8.7	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6