

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041222\
 Data File : VN071956.D
 Acq On : 12 Apr 2022 14:57
 Operator : JC\MD
 Sample : VN0412WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0412WBL01

Quant Time: Apr 13 04:44:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	666219	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1129789	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1190286	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	446442	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	358856	47.183	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.360%
35) Dibromofluoromethane	8.016	113	318923	47.972	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.940%
50) Toluene-d8	10.439	98	1417655	52.127	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.260%
62) 4-Bromofluorobenzene	12.727	95	498543	56.681	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	113.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041222\
Data File : VN071956.D
Acq On : 12 Apr 2022 14:57
Operator : JC\MD
Sample : VN0412WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0412WBL01

Quant Time: Apr 13 04:44:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

