

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041222\
 Data File : VN071960.D
 Acq On : 12 Apr 2022 17:50
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Apr 13 04:46:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	704323	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1189431	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1234098	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	473952	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	383244	47.663	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.320%
35) Dibromofluoromethane	8.022	113	342221	48.895	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.780%
50) Toluene-d8	10.439	98	1481521	51.744	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.480%
62) 4-Bromofluorobenzene	12.727	95	525405	56.739	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	113.480%
Target Compounds						
					Qvalue	
68) m/p-Xylenes	11.951	106	20216	1.258	ug/l	96
95) Naphthalene	15.504	128	61728	6.445	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041222\
Data File : VN071960.D
Acq On : 12 Apr 2022 17:50
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Apr 13 04:46:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

