

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041222\
 Data File : VN071961.D
 Acq On : 12 Apr 2022 18:14
 Operator : JC\MD
 Sample : N2349-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DEL-7

Quant Time: Apr 13 04:46:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	655286	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1118153	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1181002	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	487088	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	350552	46.860	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.720%
35) Dibromofluoromethane	8.022	113	288098	43.786	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	87.580%
50) Toluene-d8	10.439	98	1402609	52.111	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.220%
62) 4-Bromofluorobenzene	12.727	95	529048	60.775	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	121.540%
Target Compounds						
					Qvalue	
16) Acetone	4.281	43	369847	115.684	ug/l	96
20) Methylene Chloride	5.104	84	14322	2.125	ug/l #	80
25) 2-Butanone	7.333	43	175455	38.068	ug/l	99
48) Methyl methacrylate	9.539	41	25250	3.638	ug/l #	26
51) 4-Methyl-2-Pentanone	10.327	43	48308	5.023	ug/l	98
59) 2-Hexanone	11.080	43	134869	19.918	ug/l	87
95) Naphthalene	15.498	128	28088	4.973	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041222\
Data File : VN071961.D
Acq On : 12 Apr 2022 18:14
Operator : JC\MD
Sample : N2349-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DEL-7

Quant Time: Apr 13 04:46:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

