

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041223\
 Data File : VN077312.D
 Acq On : 12 Apr 2023 13:05
 Operator : JC\MD
 Sample : 02255-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COAN5

Quant Time: Apr 13 01:13:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 06 08:27:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	506153	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	963413	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	874002	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	302245	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	314300	45.434	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.860%
35) Dibromofluoromethane	8.172	113	285856	46.349	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.700%
50) Toluene-d8	10.572	98	1199312	48.304	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.600%
62) 4-Bromofluorobenzene	12.854	95	380567	46.087	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.180%
Target Compounds						
					Qvalue	
8) Diethyl Ether	3.972	74	32338	8.199	ug/l	97
16) Acetone	4.443	43	54402	21.329	ug/l	93
17) Carbon Disulfide	4.719	76	22157	1.386	ug/l #	94
20) Methylene Chloride	5.290	84	31787	4.413	ug/l #	91
37) Ethyl Acetate	7.572	43	62617	7.672	ug/l	98
43) Isopropyl Acetate	8.696	43	290945	19.320	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041223\
Data File : VN077312.D
Acq On : 12 Apr 2023 13:05
Operator : JC\MD
Sample : 02255-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
C0AN5

Quant Time: Apr 13 01:13:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 06 08:27:03 2023
Response via : Initial Calibration

