

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041224\
 Data File : VN081717.D
 Acq On : 12 Apr 2024 14:56
 Operator : JC\MD
 Sample : PB160181ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB160181ZHE#02

Manual Integrations
 APPROVED

Reviewed By : John Carlone 04/15/2024
 Supervised By : Mahesh Dadoda 04/15/2024

Quant Time: Apr 13 06:56:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	255964	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.106	114	498626	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	506950	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	191851	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	193587	59.306	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	= 118.620%		
35) Dibromofluoromethane	8.171	113	166857	53.024	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 106.040%		
50) Toluene-d8	10.571	98	643433	56.563	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	= 113.120%#		
62) 4-Bromofluorobenzene	12.853	95	207322	51.724	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery	= 103.440%		
Target Compounds						Qvalue
16) Acetone	4.430	43	56933	53.969	ug/l	98
20) Methylene Chloride	5.283	84	14297	4.137	ug/l	# 66
37) Ethyl Acetate	7.565	43	5339	1.177	ug/l	# 91
43) Isopropyl Acetate	8.688	43	322922m	42.960	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041224\
Data File : VN081717.D
Acq On : 12 Apr 2024 14:56
Operator : JC\MD
Sample : PB160181ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB160181ZHE#02

Manual Integrations
APPROVED

Reviewed By : John Carlone 04/15/2024
Supervised By : Mahesh Dadoda 04/15/2024

Quant Time: Apr 13 06:56:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

