

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041224\
 Data File : VN081722.D
 Acq On : 12 Apr 2024 16:55
 Operator : JC\MD
 Sample : P2070-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2

Quant Time: Apr 13 06:58:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	234032	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	470225	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	483161	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	179051	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	193983	64.996	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 130.000%#		
35) Dibromofluoromethane	8.165	113	152413	51.359	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.720%		
50) Toluene-d8	10.571	98	612111	57.060	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 114.120%#		
62) 4-Bromofluorobenzene	12.853	95	196412	51.961	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 103.920%		
Target Compounds						
						Qvalue
16) Acetone	4.424	43	39231	40.674	ug/l	99
20) Methylene Chloride	5.277	84	8715	2.758	ug/l #	73
43) Isopropyl Acetate	8.688	43	305037	43.032	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041224\
Data File : VN081722.D
Acq On : 12 Apr 2024 16:55
Operator : JC\MD
Sample : P2070-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2

Quant Time: Apr 13 06:58:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

