

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041224\
 Data File : VN081726.D
 Acq On : 12 Apr 2024 18:31
 Operator : JC\MD
 Sample : P2075-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 240410059-04-TRIP-BLANK

Quant Time: Apr 13 07:00:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	284548	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.106	114	563534	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	511747	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	210225	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	200272	55.190	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.380%
35) Dibromofluoromethane	8.165	113	161962	45.540	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.080%
50) Toluene-d8	10.570	98	591496	46.009	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.020%
62) 4-Bromofluorobenzene	12.853	95	186286	41.122	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	82.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041224\
Data File : VN081726.D
Acq On : 12 Apr 2024 18:31
Operator : JC\MD
Sample : P2075-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
240410059-04-TRIP-BLANK

Quant Time: Apr 13 07:00:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

