

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041319\
 Data File : VN055005.D
 Acq On : 12 Apr 2019 17:56
 Operator : JC/SP
 Sample : VN0413WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0413WBL01

Quant Time: Apr 13 03:51:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 13 03:34:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	570054	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	893882	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	718769	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	225244	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	352830	49.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.64%	
35) Dibromofluoromethane	7.59	113	284832	49.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.98%	
50) Toluene-d8	10.09	98	1047366	50.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.08%	
62) 4-Bromofluorobenzene	12.40	95	300208	42.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.16%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041319\
Data File : VN055005.D
Acq On : 12 Apr 2019 17:56
Operator : JC/SP
Sample : VN0413WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0413WBL01

Quant Time: Apr 13 03:51:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041319W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 13 03:34:24 2019
Response via : Initial Calibration

