

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041319\
 Data File : VN055008.D
 Acq On : 12 Apr 2019 19:19
 Operator : JC/SP
 Sample : K2358-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-5

Quant Time: Apr 13 04:09:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 13 03:34:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	550512	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	885891	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	715010	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	226402	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	356362	51.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.16%	
35) Dibromofluoromethane	7.59	113	282880	50.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.18%	
50) Toluene-d8	10.09	98	1032860	49.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.60%	
62) 4-Bromofluorobenzene	12.40	95	293941	42.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.14%	
Target Compounds						
16) Acetone	3.83	43	8342	4.395	ug/l	91
20) Methylene Chloride	4.55	84	60960	8.644	ug/l	97
43) Isopropyl Acetate	8.17	43	14719	1.400	ug/l	100
95) Naphthalene	15.14	128	4385	5.135	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041319\
Data File : VN055008.D
Acq On : 12 Apr 2019 19:19
Operator : JC/SP
Sample : K2358-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-5

Quant Time: Apr 13 04:09:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041319W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 13 03:34:24 2019
Response via : Initial Calibration

