

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041319\
 Data File : VN055012.D
 Acq On : 12 Apr 2019 21:05
 Operator : JC/SP
 Sample : K2204-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-04-041119

Quant Time: Apr 13 04:20:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 13 03:34:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	579453	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	902864	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	736812	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	240325	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	364740	50.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.32%	
35) Dibromofluoromethane	7.59	113	291236	50.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.20%	
50) Toluene-d8	10.09	98	1060947	50.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.38%	
62) 4-Bromofluorobenzene	12.40	95	305034	42.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.68%	
Target Compounds						
16) Acetone	3.82	43	21822	10.924	ug/l	98
20) Methylene Chloride	4.54	84	167279	25.044	ug/l	98
43) Isopropyl Acetate	8.17	43	16647	1.554	ug/l	98
95) Naphthalene	15.14	128	11218	5.783	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041319\
Data File : VN055012.D
Acq On : 12 Apr 2019 21:05
Operator : JC/SP
Sample : K2204-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-04-041119

Quant Time: Apr 13 04:20:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041319W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 13 03:34:24 2019
Response via : Initial Calibration

