

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041319\
 Data File : VN055015.D
 Acq On : 12 Apr 2019 22:25
 Operator : JC/SP
 Sample : K2358-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-7-S

Quant Time: Apr 13 04:28:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 13 03:34:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	564366	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	860200	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	712690	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	230616	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	351485	49.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.26%	
35) Dibromofluoromethane	7.59	113	276269	50.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.76%	
50) Toluene-d8	10.09	98	1029257	51.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.20%	
62) 4-Bromofluorobenzene	12.40	95	293538	43.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.54%	
Target Compounds						
16) Acetone	3.82	43	18314	9.413	ug/l	94
20) Methylene Chloride	4.54	84	1349536	218.812	ug/l	98
43) Isopropyl Acetate	8.17	43	14137	1.385	ug/l	98
95) Naphthalene	15.13	128	5575	5.249	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041319\
Data File : VN055015.D
Acq On : 12 Apr 2019 22:25
Operator : JC/SP
Sample : K2358-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-7-S

Quant Time: Apr 13 04:28:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041319W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 13 03:34:24 2019
Response via : Initial Calibration

