

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041320\
 Data File : VN060916.D
 Acq On : 13 Apr 2020 22:28
 Operator : JC/MD
 Sample : L2195-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-223

Quant Time: Apr 14 07:57:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	149161	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	254810	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	240434	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	107895	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	99511	48.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.20%	
35) Dibromofluoromethane	7.56	113	75558	49.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.08%	
50) Toluene-d8	10.08	98	317731	50.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.98%	
62) 4-Bromofluorobenzene	12.40	95	118488	51.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.74%	
Target Compounds						
16) Acetone	3.78	43	10486	14.457	ug/l	99
20) Methylene Chloride	4.52	84	3567	1.999	ug/l	90
43) Isopropyl Acetate	8.14	43	38520	8.761	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041320\
Data File : VN060916.D
Acq On : 13 Apr 2020 22:28
Operator : JC/MD
Sample : L2195-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-223

Quant Time: Apr 14 07:57:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

