

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041321\
 Data File : VN066587.D
 Acq On : 13 Apr 2021 19:04
 Operator : JC/MD
 Sample : M2021-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WGGM-21-WW-TSF-015

Quant Time: Apr 14 13:13:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N041321W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Apr 14 12:32:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.102	128	43246	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.512	114	188020	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.347	117	177630	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.944	65	121650	30.51	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	101.70%
60) 4-Bromofluorobenzene	12.345	95	57180	21.14	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	70.47%
63) Toluene-d8	10.025	98	245528	28.85	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	96.17%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041321\
Data File : VN066587.D
Acq On : 13 Apr 2021 19:04
Operator : JC/MD
Sample : M2021-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WGGM-21-WW-TSF-015

Quant Time: Apr 14 13:13:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N041321W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Apr 14 12:32:37 2021
Response via : Initial Calibration

