

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041322\
 Data File : VN071976.D
 Acq On : 13 Apr 2022 15:16
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Apr 14 02:32:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	626809	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1053169	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1087118	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	397933	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	336297	46.997	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.000%
35) Dibromofluoromethane	8.022	113	301361	48.628	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.260%
50) Toluene-d8	10.439	98	1327969	52.382	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.760%
62) 4-Bromofluorobenzene	12.727	95	446607	54.470	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	108.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041322\
Data File : VN071976.D
Acq On : 13 Apr 2022 15:16
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Apr 14 02:32:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

