

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041322\
 Data File : VN071978.D
 Acq On : 13 Apr 2022 16:04
 Operator : JC\MD
 Sample : N2349-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DEL-2

Quant Time: Apr 14 02:34:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	580302	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	987246	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1077441	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	416851	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	309805	46.764	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.520%
35) Dibromofluoromethane	8.022	113	282435	48.617	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.240%
50) Toluene-d8	10.439	98	1263398	53.163	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.320%
62) 4-Bromofluorobenzene	12.727	95	450250	58.581	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	117.160%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	50499	17.837	ug/l	95
20) Methylene Chloride	5.092	84	13936	2.335	ug/l #	88
25) 2-Butanone	7.345	43	19161	4.695	ug/l	93
43) Isopropyl Acetate	8.557	43	23767	1.609	ug/l #	87
95) Naphthalene	15.504	128	18168	4.687	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041322\
Data File : VN071978.D
Acq On : 13 Apr 2022 16:04
Operator : JC\MD
Sample : N2349-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DEL-2

Quant Time: Apr 14 02:34:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

