

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041323\
 Data File : VN077321.D
 Acq On : 13 Apr 2023 12:10
 Operator : JC\MD
 Sample : PB152070TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB152070TB

Quant Time: Apr 14 07:23:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 06 08:27:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	470075	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	904192	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	864092	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	314597	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	295002	45.917	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.840%
35) Dibromofluoromethane	8.172	113	269921	46.632	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.260%
50) Toluene-d8	10.572	98	1153284	49.493	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.980%
62) 4-Bromofluorobenzene	12.854	95	387595	50.013	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.020%
Target Compounds						
					Qvalue	
16) Acetone	4.449	43	40644	17.158	ug/l	94
20) Methylene Chloride	5.290	84	25745	3.848	ug/l	93
37) Ethyl Acetate	7.572	43	90215	11.778	ug/l	95
43) Isopropyl Acetate	8.695	43	421011	29.787	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041323\
Data File : VN077321.D
Acq On : 13 Apr 2023 12:10
Operator : JC\MD
Sample : PB152070TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB152070TB

Quant Time: Apr 14 07:23:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 06 08:27:03 2023
Response via : Initial Calibration

