

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041323\
 Data File : VN077327.D
 Acq On : 13 Apr 2023 14:36
 Operator : JC\MD
 Sample : 02292-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IDW-SO-G3-041123

Quant Time: Apr 14 07:24:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 06 08:27:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	467337	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	898796	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	842128	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.801	152	307192	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	293009	45.874	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.740%
35) Dibromofluoromethane	8.172	113	265112	46.076	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.160%
50) Toluene-d8	10.572	98	1129452	48.761	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.520%
62) 4-Bromofluorobenzene	12.854	95	379029	49.201	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.400%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	40362	17.139	ug/l	99
20) Methylene Chloride	5.284	84	38870	5.844	ug/l	96
37) Ethyl Acetate	7.566	43	87454	11.486	ug/l #	95
43) Isopropyl Acetate	8.695	43	470572	33.494	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041323\
Data File : VN077327.D
Acq On : 13 Apr 2023 14:36
Operator : JC\MD
Sample : 02292-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IDW-SO-G3-041123

Quant Time: Apr 14 07:24:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 06 08:27:03 2023
Response via : Initial Calibration

