

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041323\
 Data File : VN077329.D
 Acq On : 13 Apr 2023 15:25
 Operator : JC\MD
 Sample : PB152070ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB152070ZHE#01

Quant Time: Apr 14 07:24:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 06 08:27:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	472865	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	924072	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	872762	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	323583	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	299445	46.333	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.660%
35) Dibromofluoromethane	8.178	113	272009	45.981	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.960%
50) Toluene-d8	10.571	98	1150610	48.316	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.640%
62) 4-Bromofluorobenzene	12.854	95	371262	46.875	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.740%
Target Compounds						
					Qvalue	
8) Diethyl Ether	3.978	74	23121	6.275	ug/l	93
16) Acetone	4.448	43	69118	29.006	ug/l	97
20) Methylene Chloride	5.289	84	108307	16.094	ug/l	94
37) Ethyl Acetate	7.566	43	90190	11.521	ug/l	97
43) Isopropyl Acetate	8.695	43	516593	35.764	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041323\
Data File : VN077329.D
Acq On : 13 Apr 2023 15:25
Operator : JC\MD
Sample : PB152070ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB152070ZHE#01

Quant Time: Apr 14 07:24:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 06 08:27:03 2023
Response via : Initial Calibration

