

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041323\
 Data File : VN077331.D
 Acq On : 13 Apr 2023 16:14
 Operator : JC\MD
 Sample : PB152070ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB152070ZHE#03

Quant Time: Apr 14 07:24:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 06 08:27:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	567796	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1096040	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	1005684	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.801	152	343838	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	348218	44.872	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	89.740%		
35) Dibromofluoromethane	8.178	113	326920	46.593	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.180%		
50) Toluene-d8	10.572	98	1353426	47.915	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	95.840%		
62) 4-Bromofluorobenzene	12.854	95	408230	43.455	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	86.920%		
Target Compounds						Qvalue
8) Diethyl Ether	3.966	74	21423	4.842	ug/l	94
16) Acetone	4.449	43	72536	25.351	ug/l	97
20) Methylene Chloride	5.290	84	109403	13.539	ug/l #	92
37) Ethyl Acetate	7.572	43	92959	10.012	ug/l	96
43) Isopropyl Acetate	8.695	43	520929	30.406	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041323\
Data File : VN077331.D
Acq On : 13 Apr 2023 16:14
Operator : JC\MD
Sample : PB152070ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB152070ZHE#03

Quant Time: Apr 14 07:24:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 06 08:27:03 2023
Response via : Initial Calibration

