

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041323\
 Data File : VN077338.D
 Acq On : 13 Apr 2023 19:06
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Apr 14 07:26:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 06 08:27:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	546443	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1044071	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	898009	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	334141	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	343726	46.024	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.040%
35) Dibromofluoromethane	8.178	113	317220	47.461	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.920%
50) Toluene-d8	10.572	98	1263118	46.944	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.880%
62) 4-Bromofluorobenzene	12.854	95	375624	41.974	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.940%
Target Compounds						
86) p-Isopropyltoluene	13.736	119	37463	1.482	ug/l	Qvalue # 74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041323\
Data File : VN077338.D
Acq On : 13 Apr 2023 19:06
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Apr 14 07:26:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 06 08:27:03 2023
Response via : Initial Calibration

