

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041420\
 Data File : VN060934.D
 Acq On : 14 Apr 2020 15:30
 Operator : JC/MD
 Sample : L2229-40
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-10

Quant Time: Apr 15 09:20:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	151583	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	255796	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	240880	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	106947	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	100763	48.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.84%	
35) Dibromofluoromethane	7.56	113	75536	48.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.68%	
50) Toluene-d8	10.08	98	316946	50.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.34%	
62) 4-Bromofluorobenzene	12.40	95	119511	51.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.24%	
Target Compounds						
16) Acetone	3.78	43	20694	28.075	ug/l	96
20) Methylene Chloride	4.52	84	2825	1.558	ug/l #	93
25) 2-Butanone	6.81	43	6281	5.266	ug/l	98
43) Isopropyl Acetate	8.14	43	45649	10.342	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041420\
Data File : VN060934.D
Acq On : 14 Apr 2020 15:30
Operator : JC/MD
Sample : L2229-40
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-10

Quant Time: Apr 15 09:20:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

