

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041420\
 Data File : VN060938.D
 Acq On : 14 Apr 2020 17:00
 Operator : JC/MD
 Sample : L2230-70
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-40-41-42

Quant Time: Apr 15 09:33:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	144155	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	248172	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	233447	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	105164	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	96895	48.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.92%	
35) Dibromofluoromethane	7.56	113	72936	48.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.20%	
50) Toluene-d8	10.08	98	309079	50.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.86%	
62) 4-Bromofluorobenzene	12.39	95	116575	51.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.80%	
Target Compounds						
16) Acetone	3.78	43	13528	19.299	ug/l	99
20) Methylene Chloride	4.53	84	2858	1.657	ug/l	96
25) 2-Butanone	6.80	43	3734	3.292	ug/l #	90
43) Isopropyl Acetate	8.14	43	41729	9.745	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041420\
Data File : VN060938.D
Acq On : 14 Apr 2020 17:00
Operator : JC/MD
Sample : L2230-70
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-40-41-42

Quant Time: Apr 15 09:33:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

