

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041420\
 Data File : VN060946.D
 Acq On : 14 Apr 2020 20:01
 Operator : JC/MD
 Sample : L2230-32
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-20

Quant Time: Apr 15 09:39:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	149454	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	255509	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	241703	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	110530	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	99897	48.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.38%	
35) Dibromofluoromethane	7.56	113	75525	48.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.76%	
50) Toluene-d8	10.08	98	316693	50.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.36%	
62) 4-Bromofluorobenzene	12.40	95	120408	52.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.12%	
Target Compounds						
16) Acetone	3.78	43	15539	21.382	ug/l	97
20) Methylene Chloride	4.52	84	2934	1.641	ug/l	94
43) Isopropyl Acetate	8.14	43	44198	10.025	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041420\
Data File : VN060946.D
Acq On : 14 Apr 2020 20:01
Operator : JC/MD
Sample : L2230-32
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-20

Quant Time: Apr 15 09:39:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

