

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041422\
 Data File : VN071994.D
 Acq On : 14 Apr 2022 13:54
 Operator : JC\MD
 Sample : N2406-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK-220413002-02

Quant Time: Apr 15 01:37:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032422W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Mar 24 12:03:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.657	128	76920	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.969	114	411083	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.739	117	341471	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.434	65	158832	29.308	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	97.700%
60) 4-Bromofluorobenzene	12.727	95	110912	21.906	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	73.033%
63) Toluene-d8	10.439	98	474461	30.609	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	102.033%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041422\
Data File : VN071994.D
Acq On : 14 Apr 2022 13:54
Operator : JC\MD
Sample : N2406-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK-220413002-02

Quant Time: Apr 15 01:37:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032422W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Mar 24 12:03:44 2022
Response via : Initial Calibration

